

Chih-I Wu
01 Dec 2025



M-Fridge back up

2025/12/1 10:05:24

Device


Type of device KH 50	Serial n° OK 16.05.07540	Software version 1.06
-------------------------	-----------------------------	--------------------------

Dataset parameters

Dataset name M-Fridge back up	Type of dataset Instant
Number of points 5856	Interval 15 minute(s)
Start date 2025/10/1 10:10:00	End date 2025/12/1 09:55:00
Comments	

Channels

Channel Ch-int1 M-Fridge [°C]	Measure Temperature	Probe Ambient thermo-hygro		
Low threshold 2.0	High threshold 8.0			
Minimum 3.1	Maximum 5.2	Average 4.23	Standard 0.49	MKT 4.25





M-Fridge back up

2025/12/1 10:05:24

	Number of points	Minimum	Maximum	Average	Standard deviation	MKT
KH 50 [OK 16.05.07540] Ch-int1 M-Fridge [°C]	5856	3.1	5.2	4.23	0.49	4.25