

Chih-I Wu
01 May 2025



P-Fridge
2025/5/1 09:05:51

Device


Type of device	Serial n°	Software version
KH 50	OK 16.12.12000	1.06

Dataset parameters

Dataset name	Type of dataset
P-Fridge	Instant
Number of points	Interval
2781	15 minute(s)
Start date	End date
2025/4/2 09:55:00	2025/5/1 08:55:00
Comments	

Channels

Channel	Measure	Probe		
Ch-int1 [°C]	Temperature	Ambient thermo-hygro		
Low threshold	High threshold			
2.0	8.0			
Minimum	Maximum	Average	Standard	MKT
3.9	5.2	4.34	0.19	4.34



	Number of points	Minimum	Maximum	Average	Standard deviation	MKT
KH 50 [OK 16.12.12000]						
Ch-int1 [°C]	2781	3.9	5.2	4.34	0.19	4.34